



Tube End Construction Dimensions




	TUBE $\leq 3/4"$	TUBE $\geq 1"$
A	6"	8"
B	$1/8"$ to $3/4"$ OD	1" and up
C	1"	2"

Sanitary End Construction Dimensions



A	9" all size tubes
B	$2 1/2"$ all size tubes
C	$2" \pm 1"$ plus E
D	$1/2"$ up to 4" tri-clamp
E	$1/4"$ to 4"

Flanged Construction Dimensions



A	6	6	6	6	6	6	8	8	8	8	150 ^{RF}
B	$1/2$	$3/4$	1	2	3	4	6	8	10	12	or NPT
C	3	3	3	3	3	3	3	3	3	3	

Consult Factory

- A Non-standard length
- B 300 lb and higher
- C 6" – 200° to 250°F
10" – 250° to 550°F
18" – 550°F to 850°F



Thermal Instrument Co.

217 Sterner Mill Road, Trevoze, PA 19053

Tel: 215-355-8400 Fax: 215-355-1789

Email: office@thermalinstrument.com